


Spectrum Techniques

UCS-20

Universal Computer Spectrometer



SPECTRUM TECHNIQUES, INC.

106 Union Valley Road
Oak Ridge, TN 37830
USA

Tel. (865) 482-9937 Fax. (865) 483-0473

e-mail: support@spectrumtechniques.com

Web Site: <http://www.spectrumtechniques.com/>

Introduction	5
Screen View	7
Installation	8
Connections On Rear Panel.....	9
Installing Software	11
Uninstalling Software.....	12
Analysis Modes	13
Pulse Height Analysis	14
Multi Channel Scaling.....	15
External Multi Channel Scaling	17
Operation.....	18
Live Mode	19
File Mode	20
Amp/HV/ADC	21
High Voltage	23
Amplifier Coarse Gain	24
Amplifier Fine Gain	25
ADC Conversion Gain	26
Lower and Upper Level Discriminators.....	27
Presets.....	28
Preset Time.....	29
Preset Integral.....	30
Go, Stop and Erase.....	31
Region of Interest (ROI)	32
Functions	34
Energy Calibration	35
Isotope Matching.....	37
Strip Background.....	40
Load Spectrum	41
Load Background.....	42
Show Spectrum	43
Show Background	44
Overlay Spectra.....	45
Strip Background from Spectrum	46
Data Smoothing.....	47
Menu Bar.....	48
File.....	49
File Open	50
File Save.....	51
File Load Setup	52
File Save Setup.....	53
File Load Library	54
File Save Library.....	55
File Print.....	56
File Print Preview.....	57

File Print Setup.....	58
File Exit.....	59
Edit	60
Edit Experiment.....	61
Edit Iso Match	62
Edit Smooth Data	63
Mode.....	64
Mode Pulse Height.....	65
Mode Multi Channel Scaling	66
Mode External Multi Channel Scaling.....	67
Display	68
Display Peak Report.....	69
Display Data Report	70
Display Calibration	71
Display ROIs	72
Display Iso Match	73
Display Pixel Sizes.....	74
Settings.....	75
Settings ROIs.....	76
Settings Energy Calibrate.....	79
Settings Preset	84
Settings Amp/HV/ADC.....	85
Settings MCS	86
Settings Color.....	87
Settings Confirm Spectrum Erasure.....	89
Settings Reset All Variables To Defaults.....	90
Strip Background.....	91
Strip Background - Load Spectrum.....	92
Strip Background - Load Background	93
Strip Background - Show Spectrum.....	94
Strip Background - Show Background.....	95
Strip Background – Overlay Spectra.....	96
Strip Background – Strip Background from Spectrum	97
View	98
Tool Bar.....	99
Status Bar	100
Help	101
Contents.....	102
Using Help.....	103
About.....	104
Tool Bar.....	105
Go	106
Stop.....	107
Erase	108
Show Peak Report	109
Show Data Report	110

Amp/HV/ADC	111
Presets.....	112
ROI.....	113
Spectrum Window Sizing.....	114
Y axis Increase	115
Y axis Decrease	116
Y axis Log	117
X axis expand.....	118
X axis contract.....	119
X axis right 256 channels.....	120
X axis right 16 channels.....	121
X axis left 16 channels	122
X axis left 256 channels	123
Status Bar	124
Specifications	125
Hardware	126
Software	127
Spectrum Techniques Contact.....	128

[#]^{\$}^KIntroduction

Hardware

The Universal Computer Spectrometer offers a unique solution for nuclear spectrometry using the PC platform. A 2K ADC combined with 8K of data memory and multi channel scaling is ideally suited to scintillation spectroscopy and time related studies such as half-life decay.

Constructed in a sturdy, fully-shielded bench top enclosure with Universal Serial Bus (USB) computer interface, the multi channel analyzer contains many advanced features including computer controlled amplifier and high voltage for PM tubes, upper and lower level discriminators, on instrument data memory, and a comprehensive software package for use under Windows 98SE or higher.

The UCS-20 requires only an available USB port and is designed to work seamlessly with USB-equipped PCs. For stability and low noise operation, the unit is AC-line powered with an auto-sensing power supply for 100-250 VAC operation. An on-board microprocessor acts as the master controller and data storage device, as well as the communication link directly to the USB interface.

The integrated amplifier and high voltage are fully compatible with most standard scintillation detectors, eliminating the need for special tube bases and external modules. For ease of setup and calibration, coarse gain, fine gain, high voltage, and lower and upper level discriminator settings are controlled directly from the desktop computer. For operation with other types of detector systems such as alpha spectrometers or single photon counting, a DIRECT input is provided to bypass the scintillation amplifier and allow direct access to the ADC.

A 2048 channel ADC with de-randomizing buffer offers excellent data throughput at high counting rates with minimal dead-time losses. Conversion gain may be changed from 2048 to 1024 to 512 or 256 channels via the software. Data from the ADC is stored directly in on-board memory for autonomy and high-speed operation, freeing the host computer for other tasks.

Software

The UCS-20 produces a high-resolution, real-time live color display of spectral data with standard PC graphics running under Windows (98SE and above). Operation is intuitive using pull-down menus and function buttons for the most commonly used commands and

[#]IDH_INTRODUCTION

^{\$}Introduction

^KIntroduction

display options. The software offers full control of all features including preset live/real time, and regions-of-interest together with centroid, gross and net area calculations.

Control of the hardware amplifier, high voltage, ADC and input discriminators is also through function buttons for straightforward calibration and operation. To simplify identification of peaks, the cursor may be calibrated to read directly in energy units, using either a 2-point linear, or 3-point quadratic relationship to allow for detector non-linearity.

Spectral files may be transferred to disk for long-term storage as binary files or transferred through the clipboard in ASCII format for exporting to other programs. Stored files may be used to collect background data over a long counting period that can be subtracted on a time proportional basis from the spectral data.

To aid in the identification of nuclides, the UCS-20 contains a unique peak-labeling feature named ISOMATCH. Providing the unit is accurately energy calibrated, the user may select a nuclide from the library and the corresponding characteristic emission lines are superimposed on the spectrum along with isotope and energy information. Using this feature, users may quickly check a spectrum and visually identify the emission components for each nuclide present. The ISOMATCH libraries may be created or customized using a text editor.


When operating in multi channel scaling mode, the system may be calibrated in time units for direct readout from the cursor.

Both linear and logarithmic vertical display ranges are included which can be useful when performing decay studies

Spectrum files, can be transferred from the ICS 10 DOS program and the ICSW 16 Windows program to the UCS20 Windows version.

Most functions may be accessed by more than one method. All the functions can be accessed from the Pull Down Menu, some from the Tool Bar Buttons, or by the mouse. If the function has a Tool Bar Button, its symbol will be shown on the left hand side of the description.

#\$KScreen View



#IDH_INTROSCREENVIEW

\$Screen View

^KScreen View

#\$K Installation

System Requirements

The UCS20 instrument uses a Universal Serial Bus (USB) to communicate with a PC. The PC must have USB compatible drivers, which are available standard on Windows 98SE and later releases. The performance of the USB on older systems may affect the perceived performance of the software; so we recommend that you use a Pentium 2 or greater class of microprocessor running at 250 mHz or greater to achieve desirable performance. The executable code is less than 600 kb, but the help files are more than 5 mb; so if you have limited disk space, you may wish to leave the help files on the CD.

[Installing Software](#)

[Connections On Rear Panel](#)

[Uninstalling Software](#)

#IDH_INSTALLATION



\$Installation

KInstallation

#^{\$}^KConnections On Rear Panel

The UCS-20 is equipped with an internal Amplifier/Preamplifier allowing direct connection of a scintillation detector. If you plan to use this feature, connect the signal (anode) output of the detector tube base to the PREAMP IN connector on the rear of the instrument using a BNC cable and connect the high voltage to HIGH VOLTAGE with the MHV cable

These cables appear similar with the MHV being slightly longer than the BNC



Rear Panel

HV Output: MHV (or optional SHV) connector supplies positive 0 - 2500 volt @ 1mA maximum current to power scintillation detector. High voltage is fully regulated and computer controlled in 25 volt increments.

PREAMP IN: BNC connector for direct connection to PMT anode signal. Uses internal

#IDH_INSTCONNECTIONSONREARPANEL

^{\$}Connections On Rear Panel

^KConnections On Rear Panel

amplifier/preamp.

AMP IN: BNC connector for signal connection to external preamplifier.

DIRECT IN: BNC connector for connection to external amplifier. Bypasses internal amplifier and connects directly to ADC input. Accepts unipolar or bipolar signals 0-8 volt range.

EXT MCS: BNC connector to input positive MCS TTL pulses $> 150\text{nsec}$. 5 MHz maximum rate.

ROI OUT: BNC connector supplies a pulse out when data is acquired in a channel marked as a ROI. The pulse width is adjustable from 100usec to 25 msec by an internal pot. The pulse amplitude is adjustable from 0volt to 7.5 volts by an internal pot.

GATE IN: BNC connector for connection to external coincidence unit. A positive TTL pulse causes the ADC to accept the next pulse. The Gate pulse must be present before the peak of the input Pulse. Gate pulse width = 50nsec to 2usec.

MSB OUT: (OPTIONAL) BNC connector to output the most significant MCS bit for Mossbauer Applications.

T/COMP: Modular connector for connection to temperature compensated tube base.

HIGH VOLTAGE: MHV (OPTIONAL SHV) connector supplies positive 0 to 2500 volts to power scintillation detectors and proportional tubes. 1-Watt maximum Power available.

USB: Universal Serial Bus connector for communication with computer.

PREAMP POWER: (OPTIONAL) DB-9 connector, supplies + – 12v power for External preamplifier.

When using an external amplifier system, connect the amplifier unipolar or bipolar output to the **DIRECT IN** connector.

^{#\$K}Installing Software

Insert the included CD in your CD-ROM drive. The install program will automatically run. Follow the instructions to select the options you want to install.

The application is less than 600 kb, but the help files are greater than 5 mb; if you have limited disk space, you may not want the help file installed. You can access the help file on the CD.

The default installation directory is C:\SpecTech\UCS20

Example files:

Spectrum (*.spu) files:

Normal.spu – contains an uncalibrated spectrum of CS-137.

Calibrated.spu – contains a calibrated spectrum of CS-137.

Setup (*.sup) files:

Normal.sup – has settings used in Normal.spu.

Calibrated.sup – has settings used in Calibrated.spu.

Library (*.itm) files

Isotopes.itm – contains a list of common isotope peaks for use with the IsoMatch function.

[#]IDH_INSTINSTALLINGSOFTWARE

^{\$}Installing Software

^KInstalling Software

^{#\$K}Uninstalling Software

Go to the Windows Start Menu, select Settings, then Control Panel. On the Control Panel, select Add/Remove Software. Find UCS20, double click and follow the instructions.

[#]IDH_INSTUNINSTALLINGSOFTWARE

^{\$}Uninstalling Software

^KUninstalling Software

#^{\$K}Analysis Modes



See:

[Pulse Height Analysis](#)

[Multi Channel Scaling](#)

[External Multi Channel Scaling](#)

#IDH_MODES

\$ Analysis Modes

^KAnalysis Modes

#^{\$}K Pulse Height Analysis



This is the normal operating mode for collecting sample gamma emission spectra. The amplitude of each detector pulse is measured by the ADC and stored as an amplitude (energy) spectrum.


The X axis is scaled to the selected channels and the Y axis is scaled for the counts for each channel. Y axis counts that exceed the maximum Y axis value are 'wrapped', they no longer are scaled to the Y axis, and they appear in a different color.

#IDH_MODESPHA

^{\$}Pulse Height Analysis

^KPulse Height Analysis


#SK Multi Channel Scaling



Multi Channel Scaling mode is used for measuring time related phenomena such as half-life decay or single photon counting. The ADC is bypassed and incoming events are routed directly into memory for specific predetermined times (dwell time) and stored in sequential memory locations.

To use this mode, first acquire a spectrum using the Pulse Height Analysis mode. While acquiring the spectrum, adjust the LLD and the ULD to select the energy range of interest (for example, selecting only the 662 keV peak from Cs-137 can eliminate unwanted background and produce a superior decay curve when using a Cs-137/Ba-137 mini-generator.)

Click the Mode menu and select MCS




#IDH_MODESMCS


\$Multi Channel Scaling

KMulti Channel Scaling

Click the Settings menu and select MCS



The SetupMCS Dwell Time dialog box will appear:



Enter the Dwell Time. This time is for each memory location (channel). Remember the total pass time will be (Conversion Gain x Dwell Time).

Erase any current memory data and click start. The UCS20 will proceed to count incoming events for the selected dwell time, store the total in the first channel location, reset the counter and repeat the cycle storing each total in sequential channels.

#^KExternal Multi Channel Scaling



If you wish to use an external pulse generation system such as a coincidence circuit, it will be necessary to bypass the on-board amplifier and discriminators. Connect the external counts connector to the Ext MCS connector on the back of the instrument and select the menu item MODE then select External MCS.

When operating in this mode, the MCS input requires positive TTL signals ($>2.5\text{v}$, $>150\text{ns}$ duration).

#IDH_MODESEXTERNALMCS

\$External Multi Channel Scaling

^KExternal Multi Channel Scaling

[#]^{\$}^KOperation

[Screen View](#)

[Live Mode](#)

[File Mode](#)

[Amp/HV/ADC](#)

[Presets](#)

[Go, Stop and Erase](#)

[Region of Interest \(ROI\)](#)

[#]IDH_OPERATION

^{\$}Operation

^KOperation

#^{\$}K Live Mode



The software starts up in the Live Mode. This mode uses the USB to communicate with the instrument.


If you access the File Menu and click on Open, a new window appears which looks like the current window, except that the title bar shows 'File Mode' instead of 'Live Mode'. File mode does not communicate with the instrument and is used only to view and manipulate files that have been saved to disk.

#IDH_LIVEMODE

^{\$}Live Mode

^KLive Mode

#^{\$}K File Mode



File Mode is accessible from the live mode by going to the File Menu and clicking on Open. Select the file you want to view. Some functions, which have no use when viewing files, are disabled in the Menus and the Toolbar.

#IDH_FILEMODE

^{\$}File Mode


^KFile Mode

#\$K Amp/HV/ADC


Click the Amp/HV/ADC Button on the toolbar:



Or use the settings menu and click on Amp/HV/ADC



The Amp/HV/ADC dialog box will appear.



Adjustments may be made to the amplification, high voltage, conversion gain, low level discriminator, and the high level discriminator.

#IDH_SETTINGSAMP/HV/ADC

\$Amp/HV/ADC

KAmp/HV/ADC

Selecting OK will cause these values to be written to the UCS20 instrument.

Configuring System Parameters:

Once the program is running it will be necessary to configure the system parameters for correct operation and calibration.

Place a gamma emitting check source near the detector face. Cesium-137 (Cs-137) is a good choice. It has single peak at 662 keV.

Click on **Settings**, then click on **Amp/HV/ADC**

Set the high voltage to the voltage as listed by the detector manufacturer. As an example, set the high voltage to 600 volts, click **ON**.

***DO NOT** exceed the maximum high voltage rating of the detector, usually 1200 volts.*

Set the amplifier **COARSE GAIN** to 8, and set the **FINE GAIN** to 1 as a starting position. Click on **OK**, to set adjustments and exit the menu.

Start the data acquisition and adjust the **amplifier gain** until the 662 keV peak is approximately mid-scale.

Once the acquisition is started, you may enter the **Amp/HV/ADC** menu, make adjustments while viewing the spectrum. This will allow you to position the peak in the desired channels.

[High Voltage](#)


[Amplifier Coarse Gain](#)

[Amplifier Fine Gain](#)

[ADC Conversion Gain](#)

[Lower and Upper Level Discriminators](#)

#^{\$}K High Voltage



High Voltage may be entered directly in the text box or may be scrolled to the value desired. A warning will be issued if you attempt to input a value higher than 1200 volts, since this is the highest value that many probes can tolerate without damage. If you know that your probe can accept higher than 1200 volts, you may increase the voltage up to 2500 volts.

See also:


[Amp/HV/ADC](#)

#IDH_SETTINGSHVA

^{\$}High Voltage

^KHigh Voltage

#^{\$}K Amplifier Coarse Gain



Amplifier coarse gain may be set by clicking the radio-button next to the desired multiplier.

See also:


[Amp/HV/ADC](#)

#IDH_SETTINGSCOARSEGAIN

^{\$}Amplifier Coarse Gain

^KAmplifier Coarse Gain

#^{\$}K Amplifier Fine Gain



Amplifier fine gain may be directly entered as a multiplier value between 1.0 and 2.55.

See also:


[Amp/HV/ADC](#)

#IDH_SETTINGSFINEGAIN

^{\$}Amplifier Fine Gain

^KAmplifier Fine Gain

#^{\$}K ADC Conversion Gain



Conversion Gain represents the number of channels that will be sampled and displayed on the screen. Smaller values save as smaller files

The conversion gain default setting is 1024 channels. This is preferred for most scintillation detector applications and generally no adjustment is required.

For certain applications, such as alpha spectroscopy, it may be necessary to change this parameter to either 256, 512, or 2048 channels.

See also:


[Amp/HV/ADC](#)

#IDH_SETTINGSADCCONVERSIONGAIN

^{\$}ADC Conversion Gain

^KADC Conversion Gain

#^{\$K}Lower and Upper Level Discriminators



LLD and ULD, Lower Level Discriminator, and Upper Level Discriminator, allow the user to set a value between 0 and 106 (roughly percent) that cuts off signal before it gets to the ADC. The LLD is often used to block high-count signals in the low range that are not of interest, but require excessive time to sample. The ULD may be set up to 106. Low level discrimination (LLD) may be scrolled or directly entered.

Upper level discrimination (ULD) may be scrolled or directly entered.



An alternative method to change the LLD and ULD is to use the mouse to move the triangles under the X axis. Move the mouse over the triangle to be moved, press the left button, drag the triangle and release it where desired. The software will not allow the LLD and the ULD set outside appropriate ranges.

See also:

[Amp/HV/ADC](#)

#IDH_STARTUPLLDULD

^{\$}Lower and Upper Level Discriminators

^KLower and Upper Level Discriminators

#KPresets

Preset Time

The real time count and the live time count may be changed using the Preset function.

Click the Preset Time Button on the toolbar:




Preset Integral

The instrument can be set to stop automatically after a ROI integral.

Click the Preset Integral Button on the toolbar:



Or use the settings menu and click on Presets:



For details see:

[Preset Time](#)

[Preset Integral](#)

IDH_PRESETS

\$Presets

KPresets

#SKPreset Time




Preset Live Time provides automatic correction for counting losses caused by the system deadtime. Events which occur during the pulse processing cycle are lost to the system so the timer is automatically updated to compensate for these losses. When operating at excessively high count-rates the deadtime meter will indicate a high value and the real counting time may be more than doubled. Increasing the LLD setting can help reduce some high deadtime effects.

Preset Real Time sets the counting timer to run for actual clocktime and makes no correction for losses due to deadtime effects.

Click on **Settings**, click on **Presets**

Enter the **LIVE** or **REAL** time in the correct box. Click on **OK**, to set adjustment and exit menu.



Both the **LIVE TIME** and **REAL TIME** values are displayed on the UCS20 status bar as LT and RT. These values are saved in the file during data storage.

#IDH_OPPRESETTIME

\$Preset Time

^KPreset Time

#\$K^{Preset Integral}



Set Integral

To set an integral count it is necessary to first establish a ROI and then position the marker within the region.



Selecting an ROI and setting a value other than 0 in preset integral will cause the system to STOP counting when the total of all the channels within the ROI reaches that value.

#IDH_OPPRESETINTEGRAL

\$Preset Integral

K^{Preset Integral}

#\$K Go, Stop and Erase



Go, **Stop** and **Erase** functions are accessed with the Tool Bar buttons. Clicking on Start begins data collection. Clicking on Stop ends data collection. Clicking on Erase sets each channel's data to zero. Additionally, three 'hot key' combinations are defined as shortcuts for data acquisition functions. Pressing the appropriate action key while pressing the 'Ctrl' and 'Shift' keys will provide Go, Stop and Erase action. To Go press '*Ctrl+Shift+A*'; to Stop press '*Ctrl+Shift+S*'; and to Erase press '*Ctrl+Shift+E*'.

#IDH_OPSTARTETC


\$Go, Stop, and Erase

KGo, Stop, and Erase

#^{\$K}Region of Interest (ROI)



Region of interest (ROI) selection is an advanced feature which provides instantaneous computation of peak gross and net counts. These values may be used along with isotope decay tables and detection efficiency to calculate absolute or relative isotopic activities. ROI's must not overlap and need to be separated by at least one channel for correct area calculation. Up to 16 different ROI's are possible using the color selector from the pull down **Settings** menu.



Set ROI

To set an ROI around a peak, click the ROI button on the Tool Bar. Click the left mouse button at the beginning of the ROI, hold down the left mouse button and drag the marker to the other side of the peak, release mouse button.

Clear ROI

Move marker to the **ROI** to be cleared. Open the Setting Menu, open ROI and select Clear ROI.

Gross and Net Count

When the marker is positioned in a region of interest (ROI), the UCS20 software automatically calculates the **Gross** and **Net** area of the region. In order to minimize statistical effects at the **ROI** endpoints, a 3-point averaging technique is applied. The contents of channels $(n-1)$, (n) , and $(n+1)$ are summed and averaged to derive the content

#IDH_OPROI

^{\$}Region Of Interest

^KRegion Of Interest

of the endpoint channel for the net area computation. A linear interpolation is performed between these averaged endpoint values and counts below this interpolation are subtracted to arrive at the **Net** area of the peak. Gross counts is the sum of all channels in the **ROI**.

Position the marker in the peak of interest. The **Gross** and **Net** areas are automatically computed and displayed on the spectrum screen.

See also:


[Preset Integral](#)

#\$K Functions

[EnergyCalibrate](#)
[AutoCalibrate](#)
[Isotope Matching](#)
[Strip Background](#)
[Data Smoothing](#)

#IDH_FUNCTIONS
\$Functions
^Functions

#SK Energy Calibration




The energy calibration feature allows the marker to read directly in energy units. Two calibration functions are possible, a 2-point linear, or a 3-point quadratic fit.

In order to perform an energy calibration, it is first necessary to acquire a spectrum using known isotopes. Cs-137 together with Co-60 works well for many applications, producing gamma lines at 32 keV, 662 keV, 1173 keV and 1332 keV.

Select **2-point** or **3-point** mode and enter the calibration units to be used, (keV or MeV). Position the marker at the highest channel of the first peak and enter the peak energy value. Move the marker to the high point on the second peak to be used for the calibration, enter energy number. If a 3-point calibration is required, continue by moving the marker to the peak channel of the third peak, enter its energy and click OK. The system will now be calibrated and the marker position will read directly in energy.

To return to the channel number mode, click on **Settings**, click on **Uncalibrate**. Energy Calibration may also be selected using the right mouse button.




#IDH_FUNCENERCAL

\$Energy Calibration


KEnergy Calibration

Auto Calibrate is a convenience provided for users working in the under 1000 keV energy range. Selecting **Auto Calibrate** will initiate an acquisition sequence that attempts to determine optimum detector voltage and gain settings for a calibrated energy spectrum display. This calibration is specifically designed to place the primary peak of a Cesium-137 source at a point approximately 65% of the x-axis scale. Once located, the energy calibration coefficients are calculated to provide a two point calibration (32 and 662 keV). Use of other sources will result in erroneous calibration.




The calibration sequence can require a few minutes to complete. In the absence of an adequate source or improper cable connection, the process may not succeed. Once **Auto Calibrate** begins, an information dialog box is displayed which advises to wait for completion and provides an opportunity to cancel the calibration. Pressing of the Cancel button will result in the auto-calibration process being aborted and the high voltage setting being reset to zero. If allowed to work to completion, the gains and voltage will be left at the determined calibration settings and the spectrum scale will be displayed in energy.

#\$K Isotope Matching



Displaying Isotope Peaks

Isotope peaks may be indicated on calibrated spectra. These peaks are selected from a list provided by selecting the **Display** menu and clicking on **Iso Match**. The following dialog box will be displayed:



#IDH_FUNCISOMATCH

\$Isotope Matching

KIsotope Matching



Click the isotopes peaks in the window that you want to display. These will be highlighted to indicate selection. When your list is complete, click Okay to display them.

Editing the Iso Match list

Select the **Edit** menu and click **Iso Match**. The following dialog box will be displayed:




IsoMatch Library NOT loaded.



IsoMatch Library IS loaded

The dialog box top window labeled 'Select an existing isotope' will be empty unless an Iso Match library has been loaded. If a library has been loaded, you can select one of the entries from the window, if a library has not been loaded, or you want to add an entry to the list, enter the name in the second window titled 'or enter an isotope name'. Enter the number of peaks you want to add for this isotope in the third box labeled 'Enter number of peaks [10 max]' and click on Edit.

Further dialog boxes will appear that allow you to enter peak(s) for the isotope.




Isotope	Half-Life
Cs-137	30 years

Peak Energy (KeV)	Abundance
661.62	84.62
32.19	5.8

Buttons: Add, Delete, Edit, OK, Cancel

If you want to save the Iso Match to a disk file, open the File Menu and click on Save Library.

#\$K Strip Background



The Strip Background option is available only in the File Mode. The user may load two files (Spectrum and Background) and subtract the second file from the first. The portion subtracted is based on a time adjustment to the data in the second file. For example if the first file was measured with 100 seconds live time and the second file was measured with 200 seconds live time, then the data in the second file is divided by 2 (200 seconds / 100 seconds) before it is subtracted.

[Load Spectrum](#)

[Load Background](#)

[Show Spectrum](#)

[Show Background](#)

[Overlay Spectra](#)

[Strip Background from Spectrum](#)

Background Subtraction


This is a special case of spectrum stripping. Collect a background sample spectrum, usually for a long collection time. Load this spectrum as Background and click on 'Strip Background from Spectrum.' The live time fraction of the background is subtracted from spectrum. This provides a convenient method of removing naturally occurring background from a sample spectrum and can be very useful when working with low level environmental samples.

#IDH_STRIP

\$Strip Background

KStrip Background

#SK^LLoad Spectrum




Click on **Load Spectrum** and in the **File Dialog Box** that opens, select the spectrum you intend to have the background stripped from. For example, the spectrum may be taken for an isotope, the background may be the readings with no isotope present.

#IDH_LOADSPECTRUM1

\$^LLoad Spectrum

^KLoad Spectrum

#^{\$}K^LLoad Background




Click on **Load background** and in the **File Dialog Box** that opens, select the background you intend to strip from the first. For example, the spectrum may be taken for an isotope, the background may be readings with no isotope present.

#IDH_LOADSPECTRUM2

^{\$}Load Background

^KLoad Background

#^{\$K}Show Spectrum




Click **Show Spectrum** to view spectrum.

#IDH_SHOWSPECTRUM1

^{\$}Show Spectrum

^KShow Spectrum

#^{\$K}Show Background




Click **Show Background** to view background.

#IDH_SHOWSPECTRUM2

^{\$}Show Background

^KShow Background

#^SKOverlay Spectra




Click **Overlay Spectra** to view the spectrum and the background at the same time.

#IDH_OVERLAYSPECTRA

^SOverlay Spectra

^KOverlay Spectra

#^{\$}^KStrip Background from Spectrum




Click **Strip Background from Spectrum** to subtract the two spectra, where the background is corrected for the difference in the data collection time to give a correct proportion.

#IDH_STRIPSPECTRUM2FROM1

^{\$}Strip Background from Spectrum

^KStrip Background from Spectrum

#\$K Data Smoothing




Click **Smooth Data** to perform a 3-point averaging of the data currently being displayed. The function uses the following algorithm to average data in each channel, where ‘*n*’ is the channel number.

$$\frac{(n-1)+(n)+(n+1)}{3}.$$

#IDH_FUNCSMOOTH
\$Data Smoothing
KData Smoothing

#^{\$}K Menu Bar

Many advanced functions are possible via the pulldown menus. This section describes each operation in the sequence they appear. Black letters indicate the function is operational, gray indicates not-operational, highlighted indicates the function the be be activated.



For details:


[File](#)
[Edit](#)
[Mode](#)
[Display](#)
[Settings](#)
[Strip Background](#)
[View](#)
[Help](#)

#IDH_MENU

^{\$}Menu Bar

^KMenu Bar

#\$KFile



For details:

[File Open](#)

[File Save](#)

[File Load Setup](#)

[File Save Setup](#)

[File Load Library](#)

[File Save Library](#)

[File Print](#)

[File Print Preview](#)

[File Print Setup](#)

[File Exit](#)

#IDH_MENUFILE

\$File Menu

^KFile Menu

#^{\$}^KFile Open

File Open allows the user to open data files. A new instance of the UCS20 application in the **File Mode** is generated.

See Also:

[Live Mode](#)

[File Mode](#)

#IDH_MENUFILEOPEN

^{\$}File Open

^KFile Open

#^{\$}^KFile Save

The spectrum data and the associated setup and experiment data are saved to a *.spu file.

#IDH_MENUFILSAVE

^{\$}File Save

^KFile Save

#^{\$}^KFile Load Setup

Once the UCS20 has been setup and calibrated, all operating parameters may be stored (**Save Setup**) as a disk file for subsequent retrieval. If power has been turn off, it is a simple and convenient process to reload the setup file and quickly restore (**Load Setup**) the analyzer to its previous operating condition.

High Voltage must be turned on after loading setup !

#IDH_MENUFILELOADSETUP

^{\$}File Load Setup

^KFile Load Setup

[#]^{\$}^KFile Save Setup

Once the UCS20 has been setup and calibrated, all operating parameters may be stored (**Save Setup**) as a disk file for subsequent retrieval. If power has been turn off, it is a simple and convenient process to reload the setup file and quickly restore (**Load Setup**) the analyzer to its previous operating condition.

[#]IDH_MENUFILESVESETUP

^{\$}File Save Setup

^KFile Save Setup

[#]^{\$}^KFile Load Library

Allows loading of an Iso Match library file (*.itm).

See also:

[Isotope Matching](#)

[#]IDH_MENUFILELOADLIBRARY

^{\$}File Load Library

^KFile Load Library

[#]^{\$}^KFile Save Library

Allows saving of an Iso Match library file (*.itm).

See also:

[Isotope Matching](#)

[#]IDH_MENUFILSAVELIBRARY

^{\$}File Save Library

^KFile Save Library

#^{\$}^KFile Print

Allows the user to print the displayed spectrum.

#IDH_MENUFILEPRINT

^{\$}File Print

^KFile Print

#^{\$}^KFile Print Preview

Allows the user to preview the spectrum as it will be printed.

#IDH_MENUFILEPRINTPREVIEW

^{\$}File Print Preview

^KFile Print Preview

#^{\$}^KFile Print Setup

Allows the user to adjust the printer parameters before printing.

#IDH_MENUFILPRINTSETUP

^{\$}File Print Setup


^KFile Print Setup

#\$^KFile Exit

Closes the application.

#IDH_MENUFILEEXIT
\$File Exit
^KFile Exit

#\$K Edit



For details see:

[Edit Experiment](#)

[Edit Iso Match](#)


[Edit Smooth Data](#)

#IDH_MENUEEDIT

\$Edit Menu

KEdit Menu

#^{\$}^KEdit Experiment



The image shows a Windows-style dialog box titled "Experiment". It has a blue title bar with a close button (X) in the top right corner. The dialog contains five text input fields, each with a label above it: "Spectrum Name" (containing "Lab Test 234-432"), "Description" (containing "Determine the unknown isotope"), "Student ID" (containing "Sally 123-45-678"), "Detector Used" (containing "Acme 98765-a"), and "Comments" (containing "Peaks indicate that this is Cs-137"). At the bottom of the dialog are three buttons: "OK", "Cancel", and "Help".

Edit Experiment provides a means of inserting text into spectral file headers for referencing specific measurements. This text is saved along with data and parameter information in .SPE and .TSV files, and it is used for the Peak Report and the Data Report. The comments field saves a maximum of 50 characters.

Enter the desired text into the dialog box and click OK when finished.

#IDH_EDITEXPERIMENT

^{\$}Edit Experiment

^KEdit Experiment

#\$^KEdit Iso Match

Edit Iso Match allows the user to edit isotope data, add and delete isotopes from library.

See also:

[Isotope Matching](#)

#IDH_EDITISOMATCH

\$Edit Iso Match

^KEdit Iso Match

[#]^{\$}^KEdit Smooth Data

Edit Smooth Data allows the user to apply a three-point smoothing algorithm to the displayed data.

For details see:

[Data Smoothing](#)

[#]IDH_EDITSMOOTHDATA

^{\$}Edit Smooth Data

^KEdit Smooth Data

#\$KMode



For details see:

[Mode Pulse Height](#)

[Mode Multi Channel Scaling](#)

[Mode External MCS](#)

#IDH_MENUMODE

\$Mode

KMode

#SKMode Pulse Height



Click on **Pulse Height** to select the Pulse Height Mode.

See:


[Pulse Height Analysis](#)

#IDH_MENUMODEPULSEHEIGHT

\$Mode Pulse Height

KMode Pulse Height

#^{\$}K Mode Multi Channel Scaling



Click on **Multi Channel Scaling** to select the Multi Channel Scaling Mode.

See:

[Multi Channel Scaling](#)

#IDH_MENUMODEMCS

^{\$}Mode Multi Channel Scaling

^KMode Multi Channel Scaling

#^{\$}^KMode External Multi Channel Scaling



Click on **External MCS** to select the External Multi Channel Scaling Mode.

See:


[External Multi Channel Scaling](#)

#IDH_MENUMODEEXTMCS

^{\$}Mode External Multi Channel Scaling

^KMode External Multi Channel Scaling

^{#\$K}Display



For details see:

[Display Peak Report](#)

[Display Data Report](#)

[Display Calibration](#)

[Display ROIs](#)

[Display Iso Match](#)

[Display Pixel Size](#)

[#]IDH_MENUDISPLAY

^{\$}Display

^KDisplay

[#]^{\$}^KDisplay Peak Report

If regions of interest have been set around peaks in a spectrum, the **Peak Report** provides a convenient method of displaying peak information in tabular form. Readout will be in energy units if the energy calibration is active.

[#]IDH_MENUDISPLAYPEAKREPORT

^{\$}Display Peak Report

^KDisplay Peak Report

[#]^{\$}^KDisplay Data Report

The **Data Report** includes all hardware setting, counting parameters and spectrum data. ROI data is reported by lower and upper channels set, gross, net, FWHM, centroid, all channels and corresponding counts.

[#]IDH_MENUDISPLAYDATAREPORT

^{\$}Display Data Report

^KDisplay Data Report

[#]^{\$}^KDisplay Calibration

Display Calibration allows the user to switch between on/off in calibration mode and channel numbers or energy is displayed on the horizontal line.

[#]IDH_MENUDISPLAYCALIBRATION

^{\$} Display Calibration

^K Display Calibration

[#]^{\$}^KDisplay ROIs

Display ROIs lets the user toggle between displaying and not displaying the ROIs.


[#]IDH_MENUDISPLAYROIS
^{\$}Display ROIs
^KDisplay ROIs

[#]^{\$}^KDisplay Iso Match

Display Iso Match lets the user toggle between displaying and not displaying the Iso Match peaks.

[#]IDH_MENUDISPLAYISOMATCH
^{\$}Display Iso Match
^KDisplay Iso Match

#^{\$K}Display Pixel Sizes




Allows the user to choose between 1, 2, and 3 pixels per data point displayed. The default is 2.

#IDH_MENUDISPLAYPIXELSIZE


^{\$}Display Pixel Sizes

^KDisplay Pixel Sizes

#^{\$}K Settings



In **Preset** mode



In **MCS** mode

[ROI's](#)

[Clear ROI](#)

[Set ROI](#)

[Presets](#)

[Preset Time](#)

[Preset Integral](#)

[Amp/HV/ADC](#)

[High Voltage](#)

[Amplifier Coarse Gain](#)

[Amplifier Fine Gain](#)

[ADC Conversion Gain](#)


[Lower and Upper Level Discriminators](#)

#IDH_MENUSETTINGS

^{\$}Settings Menu

^KSettings Menu

#^{\$}K Settings ROIs



Allows the user to select the option to set or clear an ROI.

See:


[Clear ROI](#)
[Set ROI](#)

#IDH_MENUSETTINGSROIS

^{\$}Settings ROIs

^KSettings ROIs


#SK **Clear ROI**



Clicking on **Clear ROI** will clear the ROI indicated by the marker.

```
#IDH_MENUSETTINGSROICLEARROI
$Clear ROI
^KClear ROI
```

#\$KSet ROI



Clicking on **Set ROI** allows the user to set the ROI.

See also:


[Region of Interest \(ROI\)](#)

#IDH_MENUSETTINGSROISETROI

\$Set ROI

KSet ROI

#^{\$K}Settings Energy Calibrate



Opens a submenu for calibrating and uncalibrating the spectrum.

See:

[Settings Energy Uncalibration](#)

[Settings Energy Calibrate 2 Point](#)

[Settings Energy Calibrate 3 Point](#)

[Settings Energy Auto Calibrate](#)

#IDH_MENUSETTINGSENERGYCALIBRATE

^{\$}Settings Energy Calibrate

^KSettings Energy Calibrate

Settings Energy Uncalibrate

Clicking **Energy Uncalibrate** will undo energy calibration and return the spectrum to the channel mode of data display.

See:

[Energy Calibration](#)

#IDH_MENUSETTINGSENERGYUNCALIBRATE

\$Settings Energy Uncalibrate

#Settings Energy Uncalibrate

^{#\$K}Settings Energy Calibrate 2 Point

Clicking **2-Point** allows the user to calibrate the data using two points.

See:

[Energy Calibration](#)

[#]IDH_MENUSETTINGSENERGY2POINT

^{\$}Settings Energy Calibrate 2 Point

^KSettings Energy Calibrate 2 Point

^{#\$K}Settings Energy Calibrate 3 Point

Clicking **3-Point** allows the user to calibrate the data using three points.

See:

[Energy Calibration](#)


[#]IDH_MENUSETTINGSENERGY3POINT

^{\$}Settings Energy Calibrate 3 Point

^KSettings Energy Calibrate 3 Point

[#]^{\$}^KSettings Energy Auto Calibrate

Clicking **Auto Calibrate** initiates a calibration sequence. During Auto Calibration, high voltage and gain settings are automatically adjusted in sequence to determine optimum settings for energy calibration assuming detection of Cs-137 source. Once started, the operator is given an opportunity to cancel. A cancelled autocalibration does not revert to prior settings.



See:

[Energy Calibration](#)

[#]IDH_MENUSETTINGSENERGYAUTOCAL

^{\$}Settings Energy Auto Calibrate

^KSettings Energy Auto Calibrate

[#]^{\$}^KSettings Preset

Clicking **Presets** allows the user to select the Presets Dialog box.

See:

[Preset Time](#)

[Preset Integral](#)

[#]IDH_MENUSETTINGSPRESET

^{\$}Settings Preset

^KSettings Preset

[#]^{\$}^KSettings Amp/HV/ADC

Clicking **Amp/HV/ADC** allows the user to select the Amp/HV/ADC Dialog box.

See:

[Amp/HV/ADC](#)

[High Voltage](#)

[Amplifier Coarse Gain](#)

[Amplifier Fine Gain](#)

[ADC Conversion Gain](#)

[Lower and Upper Level Discriminators](#)


[#]IDH_MENUSETTINGSAMP/HV/ADC

^{\$}Settings Amp/HV/ADC

^KSettings Amp/HV/ADC

#\$^KSettings MCS

Clicking **MCS** allows the user to access the MCS Dwell Time dialog box, if the Mode is set to MCS.



Scroll to the desired Dwell Time then highlight the value selected then press OK.

See:

[Multi Channel Scaling](#)


#IDH_MENUSETTINGSMCS

\$Settings MCS

^KSettings MCS

#\$^KSettings Color

Clicking **Color** allows the user to select the **Colors Dialog Box** to set or change the colors for the ROIs, Background, .



The **Color Dialog Box** allows the user to select custom colors for ROIs, Plot Background, Plot Border and Text.


Select the item to change, click on **Change**, select the desire color, click on **OK**, returns to the **Color** screen, make additional color changes, when all color changes have been made, click on **OK**, enters data and returns to spectrum screen.

Color allows the user to select a range of **Basic Colors** or create **Custom Colors**.


#IDH_MENUSETTINGSCOLOR

\$Settings Color

^KSettings Color



Dialog Box with standard Colors.



Dialog Box expanded for custom colors.

Note:

Setup of computer graphic may affect the true color of the color selected.

[#]^{\$}^KSettings Confirm Spectrum Erasure

Unchecking **Confirm Spectrum Erasure** allows the user to bypass the confirmation of erasing current spectrum data. This option is checked each time the application is started. If the setting is unchecked, erasure of the current spectrum is immediate and cannot be undone, use it wisely. When the setting is checked, a request to erase is followed by the display of a confirmation dialog box to give an opportunity to cancel the request.


[#] IDH_CONFIRM_ERASE

^{\$} Settings Confirm Spectrum Erasure

^K Settings Confirm Spectrum Erasure

#^KSettings Reset All Variables To Defaults

Clicking **Reset All Variables To Defaults** can result in configuration settings being reset to the initialized state.




A confirmation dialog box will be presented to give an opportunity to cancel the request. Once accepted, the changes are made and cannot be undone. Click on 'Yes' to reset defaults, or on 'No' to cancel the operation.

#IDH_RESET_TO_DEFAULTS

^SSettings Reset All Variables To Defaults

^KSettings Reset All Variables To Defaults

#\$K Strip Background



Allows the user to load a spectrum and a background and then strip the background from the spectrum.

See:

[Strip Background](#)

#IDH_MENUSTRIP

\$Strip Background

KStrip Background

#\$^KStrip Background - Load Spectrum

See:

[Load Spectrum](#)

#IDH_STRIPLOADSPECTRUM1

^{\$}Strip Background – Load Spectrum

^KStrip Background – Load Spectrum

#\$^KStrip Background - Load Background

See:

[Load Background](#)

#IDH_STRIPLOADSPECTRUM2

^{\$}Strip Background – Load Background

^KStrip Background – Load Background

#\$^KStrip Background - Show Spectrum

See:

[Show Spectrum](#)

#IDH_STRIPSHOWSPECTRUM1
\$Strip Background – Show Spectrum
^KStrip Background – Show Spectrum

#\$^KStrip Background - Show Background

See:

[Show Background](#)

#IDH_STRIPSHOWSPECTRUM2

^{\$}Strip Background – Show Background

^KStrip Background – Show Background

#^{\$}^KStrip Background – Overlay Spectra

See:

[Overlay Spectra](#)

#IDH_STRIPOVERLAYSPECTRA
\$Strip Background – Overlay Spectra
^KStrip Background – Overlay Spectra

[#]^{\$}^KStrip Background – Strip Background from Spectrum

See:

[Strip Background from Spectrum](#)

[#]IDH_ASTRIPSPECTRUM2FROM1

^{\$}Strip Background – Strip Background from Spectrum

^KStrip Background – Strip Background from Spectrum

#\$KView



Allows the user to display or not display the Tool Bar and/or the Status Bar.

[Tool Bar](#)
[Status Bar](#)

#IDH_MENUVIEW
\$View
KView

#\$K Tool Bar

Allows the user to display or not display the Tool Bar.

See:

[Tool Bar](#)

#IDH_MENUVIEWTOOLBAR

\$ Tool Bar

K Tool Bar

[#]^{\$}^KStatus Bar

Allows the user to display or not display the Status Bar.

See:


[Status Bar](#)

[#]IDH_MENUVIEWSTATUSBAR

^{\$}Status Bar

^KStatus Bar

^{#\$K}Help



The **Help** menu provides a convenient operator reference for the UCS20 in the standard Windows Help format.

[#]IDH_MENUHELP

^{\$}Help Menu

^KHelp Menu

[#]^{\$}^KContents

The **Help** menu provides a convenient operator reference for the UCS20 in the standard Windows Help format.

[#]^{\$}^KUsing Help

The **Using Help** item provides the standard Windows Help on using the Windows Help system.

[#]IDH_MENUHELPUSINGHELP

^{\$}Using Help

^KUsing Help


#\$K About

Displays a splash screen with the Application name and version number.



#IDH_MENUHELPABOUT
\$ About
K About

#SK Tool Bar



[Go](#)
[Stop](#)
[Erase](#)

[Show Peak Report](#)
[Show Data Report](#)


[Amp/HV/ADC](#)
[Presets](#)

[ROI](#)

[Spectrum Window Sizing](#)
[Y axis Increase](#)
[Y axis Decrease](#)
[Y axis Log](#)
[X axis expand](#)
[X axis contract](#)
[X axis right 256 channels](#)
[X axis right 16 channels](#)
[X axis left 16 channels](#)
[X axis left 256 channels](#)

#IDH_TOOLBAR
\$Toolbar
KToolbar

#\$K **Go**




Go allows the user to start the acquire mode of a spectrum. This can also be accomplished by clicking on the green start icon on the display screen. Optionally, pressing 'A' while pressing and holding 'Ctrl' and 'Shift' keys will start acquisition.

See:

[Go, Stop and Erase](#)

#IDH_TBGO
\$Go Button
K Go Button

#\$K Stop



Stop allows the user to stop the acquire mode of a spectrum. This can also be accomplished by clicking on the red stop icon on the display. Optionally, pressing 'S' while pressing and holding the 'Ctrl' and 'Shift' keys will stop acquisition.

See:

[Go, Stop and Erase](#)

#IDH_TBSTOP

\$Stop Button

K Stop Button

#\$K Erase



Erase allows the user to erase the spectrum when in stop mode. This can also be accomplished by clicking on the Eraser Icon on the display screen. Optionally, pressing 'E' while pressing and holding the 'Ctrl' and 'Shift' keys will initiate an erase request.

Caution: If Settings-Confirm Spectrum Erasure is unchecked, erasure is immediate and final.

See:

[Go, Stop and Erase](#)

[Settings – Confirm Spectrum Erasure](#)

#IDH_TBERASE

\$Erase Button

K Erase Button

[#]^{\$}^KShow Peak Report



If regions of interest have been set around peaks in a spectrum, the **Peak Report** provides a convenient method of displaying peak information in tabular form. Readout will be in energy units if the energy calibration is active.

[#]IDH_TBSHOWPEAK

^{\$}Show Peak Button

^KShow Peak Button

#\$K Show Data Report



Data Report includes all hardware setting, counting parameters and spectrum data. ROI data is reported by lower and upper channels set, gross, net, FWHM, centroid, all channels and corresponding counts.

#IDH_TBSHOWDATA

\$Show Data Button

KShow Data Button

#^{\$}K Amp/HV/ADC



Allows the user to select the **Amp/HV/ADC** Dialog box.

See:

[Amp/HV/ADC](#)

[High Voltage](#)

[Amplifier Coarse Gain](#)

[Amplifier Fine Gain](#)

[ADC Conversion Gain](#)

[Lower and Upper Level Discriminators](#)

#IDH_TBAMPHVADC

\$AMP/HV/ADC Button

^KAMP/HV/ADC Button

[#]^{\$}^KPresets



Allows the user to select the **Presets** Dialog box.

See:

[Preset Time](#)

[Preset Integral](#)

[#]IDH_TBPRESETS

^{\$}Presets Button

^KPresets Button

^{#\$K}ROI



Allows the user to set an ROI.

See:

[Region Of Interest](#)

[#]IDH_TBROI

^{\$}Region Of Interest Button

^K Region Of Interest Button

[#]^{\$}^KSpectrum Window Sizing



These buttons allow the user to set the size of the X and Y axes.

If a button is grayed, its function is not available for use.

[#]IDH_SPECTRUMWINDOW

^{\$}Spectrum Window

^KSpectrum Window

#^{\$}^KY axis Increase



Clicking the Y axis increase button doubles the magnitude of the Y axis.

#IDH_TBYINC

^{\$}Y axis increase Button

^K Y axis increase Button

#^{\$}K Y axis Decrease



Clicking the Y axis decrease button halves the magnitude of the Y axis.

#IDH_TBYDEC

^{\$}Y axis decrease Button

^KY axis decrease Button

#^{\$}K Y axis Log



Clicking the Y axis Log button sets the Y axis to a logarithmic scale with a maximum of 16 meg.

#IDH_TBYLOG
^{\$}Y axis Log Button
^KY axis Log Button

#^{\$}K X axis expand



Clicking the X axis expand button increases the span of the X axis, up to the maximum set by the Conversion Gain.

#IDH_TBXEXP

^{\$}X axis expands Button

^KX axis expands Button

#^{\$}K X axis contract



Clicking the X axis contract button decreases the span of the X axis.

#IDH_TBXCONT

^{\$}X axis contract Button

^KX axis contract Button

#^{\$}K X axis right 256 channels



Active if there are 256 channels to the right of the displayed X axis span. Clicking this button will shift The X axis span 256 channels to the right.

#IDH_TBXR256

^{\$}X axis right 256 channels Button

^KX axis right 256 channels Button

#^{\$}K X axis right 16 channels



Active if there are 16 channels to the right of the displayed X axis span. Clicking this button will shift The X axis span 16 channels to the right.

#IDH_TBXR16

^{\$}X axis right 16 channels Button

^KX axis right 16 channels Button

#^{\$}K X axis left 16 channels



Active if there are 16 channels to the left of the displayed X axis span. Clicking this button will shift The X axis span 16 channels to the left.

#IDH_TBXL16

^{\$}X axis left 16 channels Button

^KX axis left 16 channels Button

#^{\$}K X axis left 256 channels




Active if there are 256 channels to the left of the displayed X axis span. Clicking this button will shift The X axis span 256 channels to the left.

#IDH_TBXL256

^{\$}X axis left 256 Button

^KUSC20 Topics

####SK Status Bar



The status bar at the bottom of the window displays a context sensitive help message on the left side, the high voltage, live time, real time, and selected device on the right side.

#IDH_STATUSBAR

#IDH_SBHV

#IDH_SBLT

#IDH_SBRT

\$Status Bar Button

KStatus Bar Button

[#]^{\$}^KSpecifications

[Hardware](#)
[Software](#)

[#]IDH_SPECIFICATIONS
^{\$}Specifications
^KSpecifications

[#]^{\$}^KHardware

- Physical:** Single instrument includes preamplifier, amplifier, detector high voltage, 2048-channel multichannel analyzer with data memory, LLD and ULD. Fully compatible with many scintillation detectors and commercial tube bases.
- Amplifier:** On-board combination preamplifier/amplifier for use with scintillation detectors and PMT's. Computer controlled coarse and fine gain from (x) 1 to (x) 160.
- ADC:** Wilkinson type with 80 MHz clock and computer selected conversion gain of 2048, 1024, 512, or 256 channels. Direct input accepts pulse peaking times of 1 μ sec. to 10 μ sec. Includes dead-time correction when used in Live-time mode.
- LLD & ULD:** Independently computer controlled in 4-channel increments over entire input range. Operates prior to ADC for reduced system dead time.
- Modes:** MCA for pulse height analysis, or MCS for half-life decay or other time related studies.
- Timers:** Real-time or Live-time operation selectable in 1-second increments for PHA, or dwell times from 10 msec. to 600 seconds per channel in MCS mode.
- Data Memory:** On-board static RAM, 2048 channel (x) 3 Bytes for data, plus region-of-interest flag.
- Deadtime:** System dead-time is computed and displayed on screen during acquisition.
- Power:** AC line powered with an auto sensing power supply for 100 – 250 VAC, 10 watts total.

Connections On Rear Panel:

[Connections On Rear Panel](#)

[#]IDH_SPECHW

^{\$}Hardware Button

^KHardware Button


[#]^{\$}^KSoftware


System:	Operates under Microsoft Windows® 98, Windows® 2000, Windows® XP with 600 kb for the executable and 6 meg for the optional help file available disk space, 32 meg RAM (48 recommended), and compatible mouse.
Display:	VGA or SVGA color graphics (800x600 recommended. Vertical scale adjusts from 32 to 16M and LOG display. Horizontal 2048 channels with reduction down to 256 channels.
Time Mode:	Preset live-time or preset real-time selection. Both times are recorded and displayed.
ROI'S:	Multiple Regions-of Interest using color coding.
Integral:	When cursor is in ROI, computes gross area, net area with end point averaging, centroid and FWHM.
Energy Cal:	2-point linear or 3-point quadratic converts cursor position reading directly to energy units. (Time units in MCS mode.)
Strip:	Subtracts channel-by-channel time normalized data stored in File Mode. Overlays two saved spectra in File Mode
Smooth:	3-point smoothing of selected displayed data.
Control:	Software control of High Voltage, Amplifier Gain, Lower and Upper level discriminators, and ADC Conversion Gain.
File:	Save or load data file and header information in binary, or spreadsheet compatible formats.
Print:	Prints current screen display to Windows printer.
ROI File:	Saves ROI data, centroid, FWHM, gross and net integrals, and header information in spreadsheet compatible format.
IsoMatch:	Isotope library text file with peak markers and labeling for overlaying on spectrum for quick isotope identification. Library may be edited from Edit, IsoMatch pull down menu.


[#]IDH_SPECSW

^{\$}Software Button


^KSoftware Button







LAST ECN R99 15K>>>7.5K ECN 01-18-02 R-125 & R-128 SWAP		
Title UCS-20 ANALOG CIRCUITS		
Size D	Number	Rev 1
Date 11-14-00	(REV1 03-22-01)	Drawn by
Filename UCS-20.S02	Sheet 2	of 3



Title			UCS-20 POWER SUPPLIES		
Size	Number		Rev		
C			1		
Date	05-05-00	(LAST 02-08-02)	Drawn by	R.S.S	
Filename	UCS-20.S03		Sheet	3	of 3

[#]^{\$}^KSpectrum Techniques Contact

SPECTRUM TECHNIQUES, INC.

106 Union Valley Road

Oak Ridge, TN 37830

USA.

Tel. (865) 482-9937 Fax. (865) 483-0473

e-mail. spectech@esper.com

Web Site. <http://www.spectrumtechniques.com>

[#]IDH_SPECTECHCONTACT

^{\$}Spectrum Techniques Contact

^KSpectrum Techniques Contact